

Denton DCP-1 critical point dryer, PAA A059

Specification

- for ethanol dehydrated samples, with liquid CO₂ as the transitional fluid
- pressure chamber is 1-1/8" diameter by 2-5/8" deep (28 mm by 66 mm)