

lyophilizers, PAA A079

VirTis BT6KEL-85 freeze dryer

- condenser temperature: -80 °C
- vacuum control (VacSet) range: above 5 mtorr or 0.67 Pa
- specimen can use the condenser chamber, or rack or flask with the manifold

Labconco FreeZone 2.5 L Benchtop freeze dryer

- condenser temperature: -50 °C
- vacuum set point range: 165 Pa to 0.2 Pa
- specimen can use the condenser chamber or flask with the manifold